

DAQ session

Introduction & Short status

Vincent Boudry
LLR, École polytechnique

Calice eDAQ
LLR
09 /02/2011

TB DAQ modes

Single Event + Ext. Trig

- ▶ External trigger (from hodoscope or calibration system) = HOLD
 - ◆ Stop Acq, Hold analog data + sampling, Start Acq
- ▶ Noise & Beam condition safe (only 1 evt per trigger)


Single Event + auto-Trig NOW USED IN SDHCAL TB

- ▶ **External trigger** (hodoscope) → DIF
 - ◆ Stop Acq, ReadOut (last evt ~ triggered one), Start Acq
(AHCAL “validation mode” identical: no external trigger within memory depth = lost data)
- ▶ **Data sync** (for Event building)
 - ◆ On synchronized BC ID → **need for a SYNC @ MClk (100- 400 ns)**
 - ◆ On trigger timestamp (e.g. On DIF Timecounter on last internal trigger to ext. trigger):
$$\text{time} = (\text{BC-LastBC}) \times \tau_{\text{BC}} - \text{DiffCounter} \times \tau_{\text{DiffCounter}}$$
 - ◆ **BUT: for the AHCAL/Spiroc:** the TDC signal needs a SYNC of the clocks **±1ns**
- ▶ **Rems:** RAMfull → Reset of SLAB with BUSY

ILC like

- ▶ StartAcq on Start-of-Spill signal (-δt)
- ▶ StopAcq & Readout on End-Of-Spill or RAMfull or a Given # Beam Trigger

CALICE DAQ2 scheme



Vincent.Boudry@in2p3.fr

- LDA-DIF on HDMI (Config, Control, Data, Clock, Trig, Busy, Sync)
- Clock, Trig, Busy & Sync on HDMI (compatible LDA-DIF)
- Optique (alt. Cable) GigE
- Debug USB
- External Trigger

ODR = Off Detector Receiver
LDA = Link Data Agregator

DCC = Data Concentrator Card
DIF = Detetcor InterFace

CCC = Clock & Control Card

HW availability

Card	#Avail	#Tested	#OK	Remark
PC	6	6	6	OS needs upgrade
ODR	10	4	4	(commercial board: no expected default)
LDA	25	22	17	
HDMI Mezzanines	30	24	13	4 have faulty connectors and are being repaired. Not all cards have 10 conn. working
GEth mezzanines	25+5	25	20	2 can easily be recovered
CCC Adapter	25	17	16	Limits # of installations
CCC	10	10	10	term adaptation maybe be needed
DCC	2+20	22	21	1 faulty channel on 1 card; 1 burned to be repaired
ECAL DIF	29	29	29	equipement for ~20 additional ones avail.
SDHCAL DIF	190	190	185	2 dead ; mods needed for HR2 (ok for HR2b)
AHCAL DIF	4*			*Being produced

All basic HW avail.

Limits # of installations


term adaptation maybe be needed

News from Matthew Wing:

- Remaining ECAL DIFs: We have **produced 20** (*ver. 2*), with 10 each in system tests at LLR and UCL. The **remaining 20** will be produced soon. The people at Cambridge have another job to do at the moment, so if you want to give us a time-scale, that's fine, otherwise they will be done soon.
- CCCs. We are in the process of sending two more cards to Guillaume. This will mean that 9 have been distributed and we'll keep one at UCL.
- HDMI cables. We recently bought 50 cables and have distributed many of them. These were pretty cheap and we got 2m cables. The specs can be found at : <http://www.toby.co.uk/content/catalogue/products.aspx?series=TRB-HH-C-xM>
 - ▶ **rem: 40 @ LLR ; not halogen free.**
- The firmware for the trigger and busy and for the HDMI connections is done, as per Matt's e-mails. We consider that we have done as much firmware as our time will allow and are concentrating on getting the various bits of hardware tested and sent out.
- The hardware page in the twiki should be reasonably up-to-date, although we are checking this. And as said above we are going through all the loose ends, doing tests and will distribute all components. I assume we will send them all to LLR.
 - ▶ **To everyone: please this page up-to-date**
<https://twiki.cern.ch/twiki/bin/view/CALICE/HardwareList>

Cables

- CERN requires halogen free cables
 - ▶ “IS23 does apply to above-ground installations and experiments.”
- On shelf: only for HiFi freaks (or Pigeons):
 - ▶ beautiful 100€ apiece 5m-long shielded HDMI cable
- 1 reasonable offer (by Franck)
 - ▶ On demand PolyEthylene LSOH coating
 - ▶ ~ 25€/cable (5m long, Ø 8.5mm) for 200+ cables.
 - ▶ pbm: 12 weeks delays (8wks by plane, +30%)
 - ▶ ~ enough funds on ANR to buy for the m³ SDHCAL (150 needed)
 - ◆ Urgent : 12 weeks delay due to boat shipping from China
 - ◆ Other demands being surveyed:
 - μMegas (~30 ?)
 - AHCAL (50) and ECAL (30)
 - + 10% spares (enough ?) → 260-275



Check F. Davin presentation

Order today

FW Performance Map

F. Gastaldi (LLR) + M. Warren (UCL)

	LDA	DCC	DIF's
Ethernet	✓ at full speed		
CCC	Clk; Trig; Busy	Clk; Trig; Busy	Clk; Trig; Busy
Nlinks up	10 MUX OK @ 25 MHz Still instable for >6 @ 50MHz	9	1
Fast Commands	✓	✓	✓
Block transfert (Config loading)	✓	✓	✓
Data	✓ (< 50 MHz)	✓ (< 50 MHz)	✓ (<50 MHz)
ROC			Structure ✓ Adapt SDHCAL USB Config loading ✓ Rest on going

Generic code for all DIFs

G. Vouters (LAPP)+ R. Cornat (LLR)




Integration tests

- Working Bench test @ UCL, LLR, Cambridge and **now LAPP* and IPNL**
- Whole chain established :
DAQ PC with ODR ↔ LDA ↔ DIF and CCC source
- Multiple 10 DIF ↔ LDA links established @ high speed
- FastTrig and Busy signals functional.




ECAL FEV7

@LLR



RPC SDHCAL ASU
Somewhat "limiting factor"

- Interactive hardware test software (GUI)
 - ▶ Each HW test easily scriptable: simple user-friendly python API: each function defined ↔ 1 graphical pane with “Run” button
 - ▶ Available to anyone working with USB/RS/Ethernet devices
- C libraries implementing the complete **DIF Task force** protocole → API



```
File Edit Options Buffers Tools Python Help

def send_FC_DCC_get_status(INT0x_lda_out_mask = 0x8):
    """Send FCMD K28.3/D15.0 (aka. 7C/D15.0, DCC get status) and
    print out the whole DCC register page"""
    comma = commons.encode_8b10b_kd(28, 3)
    data = commons.encode_8b10b_kd(15, 0)
    ans = LDA.do_lda_send_fastcmd(INT0x_lda_out_mask, comma, data)
    calicediag.GUI.set_statusbar_message("Get_Status FCMD sent")

    return _unpack_DCC_get_status_page(ans[16:]) is not False

--:--- DCC.py 47% (189,0) SVN-1428 (Python)-----
calicediag.register_action(DCC.send_FC_DCC_reset)
calicediag.register_action(DCC.send_FC_DCC_get_status)
calicediag.register_action(DCC.send_FC_DCC_init_links)

calicediag.register_action(DCC.send_BT_DCC_get_status)
calicediag.register_action(DCC.send_BT_DCC_config_tx_rx)
calicediag.register_action(DCC.send_BT_DCC_start_RTT)
calicediag.register_action(DCC.send_BT_DCC_stop_RTT)
calicediag.register_action(DCC.send_BT_DCC_relock_DCM)
calicediag.register_action(DCC.send_BT_DCC_register_blob)


--:--- GUI_DCC.py Bot (7,0) SVN-1428 (Python)-----
```

<https://svn.in2p3.fr/calice/online-sw/trunk/pyserdiag/>

SW status

- Missing critical elements
 - ▶ Configuration DB (being worked on)
 - ▶ DAQ2 interface ↔ XDAQ being worked on
- Missing ancillaries
 - ▶ Semi-automatic noisy channels spotting & correcting (monitoring)
 - ▶ Clean Slow control
 - ▶ interface to CondDB;
 - ▶ event display : DRUID on LCIO file
 - ▶ interface to the GRID
 - ▶ interface to the machine (⇒ in AIDA WP8.6.2)
Code exists in DAQv1

Implemented



Pending questions for test beams

- November test with CALICE ECAL ϕ al prototype → requires big adaptations
 - ▶ readout of CRC (VME 9U)... in xDAQ or in DAQv1⊕ DAQv2
 - ◆ Exp. from WHCAL + μ Megas for HW sync → Idem or use a EUDAQ TLU ?
 - ◆ Much work for the SW: Paul not necessarily available
- Slow Control
 - ▶ DIM
 - ◆ t^0 measurement
 - ◆ pressure meas
 - ◆ HV
 - ◆ Humidity
- Machine interface
 - ▶ code available in DAQv1 (Sven Karstensen from DESY)
 - ▶ probably to be improved (slow) ↔ part of AIDA task
 - ▶ Readout of machine events → BIF card.
- Performances for μ Megas : 4 partitions × no rate limitations → Highest BW ?

Pending questions for test benches

- Adaptation of configuration files (HR → Spiroc & Skyroc; μ ROC should be transparent)
- Adaptation of xDAQ online loops (threshold scans, calibration signal, ...)
 - ▶ auto generation of config files
- Quick online analysis
 - ▶ Adaptation of DHCAL Marlin tools → pseudo online analysis.
- Priorities ?



Conclusion

- All HW is there and working “well enough”
 - ▶ Dispatching started (LLR, IPNL, LAPP, DESY)
 - ▶ still some non-perfect parts (esp. LDA)
 - ◆ extensive test of all LDA to be done
- Last two critical missing interfaces in good progress (f early january)
 - ▶ ROC handling through HDMI
 - ◆ allow for config loading
 - ▶ DAQv2 HW ↔ xDAQ driver
 - ◆ allow for heavy load testing of DAQ SW.
- Still much work needed on the SW side (check Laurent's talk)
 - ▶ Slow control to be developed (IPNL & LLR)
 - ◆ // development on ECAL now possible when SW adapted...

Spares

Clock and Control Card


- Developed at UCL (M. Warren, M. Postranecky)
- Distributes on 8 channels (HDMI, SMAs, NIM, ...) via dedicated circuitry for **low jitter**
 - Int | ext clock
 - Fast Signal (Trigger | Sync)
- Sums-up BUSY
- Performs Trigger logics
 - CPLD
- Was used as DIF-Master (dev^t of LAPP)
 - Aka also sending hard-coded commands to DIF directly
 - Standalone tests with USB readout



Software: XDAQ framework

- dev^{ts} started @IPNL for electronics test using XDAQ in 2008
 - ▶ Ch. Combaret (IPNL)
 - ▶ Gained (a lot of) impulsion with involvment of L. Mirabito (resp. of DAQ SW for CMS tracker)
- Ran for ≥ 1 year in TB, Cosmics & Electronics test
 - ▶ USB readout
 - ▶ Interface to old LabView program
- Recent development
 - ▶ Writing of LCIO data in RAW format
 - ▶ versatile online analysis framework (root histos)

→ Marlin Based



IPN Lyon

Performances

- Rather low demands in term of bandwidth (but \gg @ ILC for same vol.)
 - SDHCAL : $\sim 20\text{MB/s}$ in Spill
 - ECAL: $\sim 100\text{MB/s}$
 - AHCAL: $\sim 300\text{MB/s}$
- Data limited by ASICs readout
 - Modes:
 - test beam single event
 - Test beam burst (\approx ILC-like mode)
- Some code (System C, by D. Decotigny) exists for simulation of full chain
- Full scale test scheduled (L. Mirabito)

DAQv2 data flux

N DIF/LDA	N DIF/DCC	LDA-DIF Dclk [MHz]	LDA-DIF FLUX [MB/s]	LDA Dclk [MHz]	LDA FLUX [MB/s]	ODR FLUX [MB/s]	Disk Flux [MB/s]
10	9	50	6.25	1000	125	1000	170

Detector	DHCAL	Evt Size	Mem Size	ASIC Dclk [MHz]	ASIC FLUX [MB/s]
		20 B	128	2.5	0.31

from LC-DET-2004-029

Mode	Calib/Noise Single	Calib/noise Burst	TB Single	TB Burst	Demo	Occupancy for 100 GeV π in TB evts
N ASIC/DIF	48	48	4.8	4.8	4.8	4.8
σ (NASIC)	0	0	2.6	2.6	2.6	2.6
Touched DIF/pla	3	3	1	1	1	1
			$+3\sigma$	$\sqrt{\text{MemSize}}$		5.49

Parameters codes	Hardware (~fixed)	DAQ (achievable)	Physics (occupancies)

ASIC	20 B	2 560 B	20 B	2 560 B	2 560 B
<i>R/O time 1</i>	64 μ s	8 192 μ s	64 μ s	8 192 μ s	8 192 μ s
<i>R/O time ALL</i>	3 072 μ s	393 216 μ s	307 μ s	39 322 μ s	39 322 μ s
DIF	960 B	122 880 B	96 B	12 288 B	12 288 B
<i>R/O time</i>	154 μ s	19 661 μ s	15 μ s	1 966 μ s	1 966 μ s
LDA w/o DCC	9 600 B	1228 800 B	320 B	40 960 B	40 960 B
<i>R/O time</i>	77 μ s	9,830 μ s	3 μ s	328 μ s	328 μ s
DCC	8,640 B	1,105,920 B	288 B	36,864 B	36,864 B
<i>R/O time</i>	1 382 μ s	176 947 μ s	46 μ s	5 898 μ s	5 898 μ s
LDA w/ DCC	86,400 B	11,059,200 B	2,880 B	368,640 B	368,640 B
<i>R/O time</i>	691 μ s	88 474 μ s	23 μ s	2 949 μ s	2 949 μ s
ODR	172,800 B	22,118,400 B	5,760 B	737,280 B	737,280 B
1000MB/s	173 μ s	22 118 μ s	6 μ s	737 μ s	737 μ s
Disk	172,800 B	22,118,400 B	5,760 B	737,280 B	737,280 B
170MB/s	1 016 μ s	130 108 μ s	34 μ s	4 337 μ s	4 337 μ s
Max R/O time	3 072 μ s	393 216 μ s	307 μ s	39 322 μ s	39 322 μ s
Min Freq	0.33 kHz	0.00 kHz	3.26 kHz	0.03 kHz	0.03 kHz
Min. evts Freq		0.33 kHz		3.26 kHz	3.26 kHz

19MB/s

Reliability tests

Stress tests using pseudo-random generator

- 9×DIF → 1×DCC → 1×LDA → PC
 - ▶ 9 DIFs (ECAL & SDHCAL) generate pseudo random data
- Results
 - ▶ Direction DIF → LDA ✓
 - ▶ Maximum DCC → LDA link occupancy (40Mbps) ✓
 - ▶ Up to 5.6 TB transferred (2 weeks), no error

End-to-end test: FIFO write/read


- PC ↔ 1×LDA ↔ 1×DCC ↔ 1×DIF
 - ▶ Tests both fast-commands and block transfer “read” requests
- PC ↔ LDA Ethernet OK

ROC config loading & checking ✓

To be done


- FW
 - ▶ Stress tests of LDA (needed: HW sometimes uneven)
 - ▶ SW: recognition & recovery of HW failures (plug-in / out)
 - ▶ Measurement of Clock & trigger dispersion
- HW support for CCC & LDA
 - ▶ LDA mechanics “unconventional” and fragile

(longueur = 22.5 cm ; Largeur = 14.5 cm ; hauteur = 6 cm)



→ M. Anduze

F. Gastaldi, N. Roche



- Man power available, no show stopper...

FW status

- DCC : not much to say
 - ▶ 1 card with 1 channel to be repaired
 - ▶ 1 card with P/S burned last week
 - ▶ Distribution: 1 @ LAPP, 1 @ IPNL
- LDA
 - ▶ Still some problems: not all channels working on all cards (HDMI mezzanine “tricky”)
 - ◆ Works @ 25 MHz
 - ◆ Unstable @ 50 MHz with triggers (passing nearby) on some channels
 - ▶ Multiplexing working fine on 6 channels (long running),
 - ◆ somewhat OK on 7 (several days),
 - ◆ problematic on 8-10 channels (blocking of card @ high data rate) but OK with pauses in data flux (as it should be the case for TB).
- CCC
 - ▶ ~OK
 - ◆ with 1 μ s-long signals (due to AC coupling)
 - ◆ Jitter on all chain to be measured... (no pbm for SDHCAL)

DIF FW

- FW still in building:
 - ▶ Done:
 - ◆ Data pseudo-random pattern sending
 - ◆ Config reception, storing and echoing
 - ◆ Config of ROC chips (on RPC ASU, ECAL FEV7) with X-Check yet
 - ◆ Soft Reset
 - ◆ Trigger/Busy echo
 - ▶ Existing but to be integrated & tested
 - ◆ Readout
 - ◆ State machine for Acquisition
 - ⇒ BUSY output...
 - ▶ To be done
 - ◆ Data formatting: on-going
 - ◆ Sync mechanism (5 MHz clock out of 50 MHz on trigger signal)

Beam InterFace card

Basis:

- CALICE chips use auto-trigger
 - ▶ Readout can be triggered by single event using external trigger (e.g. beam hodoscope)
 - “Single event” mode
 - History of Chip is usable (e.g. in case of selective ext. trigger)
 - ▶ Readout triggered by environmental internal or external trigger
 - ◆ Chip full
 - ◆ ILC-like mode (end-of-spill)
- Require some device to readout the beam line parameters
 - ▶ Scintillators; Cherenkov PM (coding of CEDAR bits)
 - ▶ Time of event (⇒ rec for wire chambers) within a 5 MHz clock period

Implementation

- 2 solutions
 - ▶ Add-hoc card for interfaces with a CALICE ROC (SPIROC ?) + 1 DIF
 - ▶ Small adaptation (buffers) card on a DIF + “simulation” of a digital ROC in the FPGA
 - ◆ Part of the coding can be “tricky”
- Both offer full compatibility with CALICE DIF for the DAQv2.
- To be implemented for 2nd version of CALICE beam test
- One of the tasks of AIDA (WP8.6.2)
 - ▶ For “standalone” CALICE tests
 - ▶ Functionalities ⇒ in JRA1 TLU

Use of sub-ns TDC for CERN wire chambers until then ?